

## **SUMMARY OF INTERVIEW**

### Attendees, Date and Type of Interview

Applicants thank Examiner Robert Nasser for the telephonic interview with Applicants' representative, Laura Johnson, on May 26, 2010 to discuss the pending application.

### Exhibits and/or Demonstrations

None.

### Identification of Claims Discussed

Claims 88, 96, and proposed new claims.

### Identification of Prior Art Discussed

None.

### Proposed Amendments, Principal Arguments and Other Matters, Results of Interviews

Applicants' representative proposed amending the claims to replace the term "sensing mechanism" with the term "sensor." Applicants' representative also proposed a new independent claim which recites that the device is capable of an accuracy level in which at least 95% of glucose concentration values measured by the device are within 25% of corresponding values determined by analysis of blood over a time period of at least 2 days. The Examiner agreed that the proposed amendments and the new claims appeared to be supported in the specification as filed, but further consideration would be required.